

Intended use

Specimen receptacle for the containment of blood samples for laboratory analysis.

Intended users

Professional use only.

Additive

Empty.

Handling

Single in vitro use only.

Use for the intended purpose only.

Use appropriate personal protective equipment when handling.

Contamination risk if contents leak – ensure lids and caps are applied correctly.

Do not use if packaging is damaged or product is not sealed.

Only re-cap with correct caps.













Dispose of using appropriate personal protective equipment and act according to applicable regulations.

Tubes maximum centrifugation: 3000 x g for 30 minutes.

Storage

Store in a dry place at room temperatures. Tubes can be refrigerated but it is not necessary to do so.

Keep dry and away from direct sunlight or sources of heat.

	Manufacturer		Indicates the date after which the product should not be used		Single use only
	Authorized representative in the European Union		Manufacturer's batch code		<i>in vitro</i> diagnostic medical device
	Consult this instruction for use		Manufacturer's catalogue number		CE marked (device is in conformity with EU requirements)
	Do not use if package is damaged		Keep away from sunlight		Keep dry